eXact® Strip Micro 7+ Reagent Reorder Information

eXact® Strip Micro (4mL) Reagent Specifications - For use with eXact® Micro 7+, Part no. 486691

No.	PARAMETER	PART NO.	# OF TESTS	DETECTION RANGE	CHEMISTRY
	eXact® Micro Carrying Case w/ foam*	486001	N/A	N/A	N/A
	Dilution Kit	487200	N/A	N/A	N/A
	Reference Standard	486602	15	N/A	N/A
	Mini Dilution Kit II	487202	N/A	N/A	N/A
1	Alkalinity, Total (as $CaCO_3$)*	486641	100	10 - 200 ppm	Alizarin Red S + Citrate
2	Bromine, Total*	486644	100	0.05 - 13 ppm	DPD
3	Hardness, Total High (as $CaCO_3$)*	486656	50	30 - 550 ppm	Phthalein Purple
4	Chlorine, Free*	486637	100	0.05 - 11 ppm	DPD
	Chlorine, Free	484051	100 Foils	0.05 - 11 ppm	DPD
5	Chlorine, High Range*	486672	50	1.2 - 250 ppm	KI + Buffer
6	Chlorine, Combined* **	486638	100	0.05 - 11 ppm	KI
	Chlorine, Combined* **	484053	100 Foils	0.05 - 11 ppm	KI
7	Chlorine, Total*	486670	100	0.05 - 11 ppm	DPD + KI
	Chlorine, Total	484054	100 Foils	0.05 - 11 ppm	DPD + KI
8	Copper (as Cu^{+2})*	486632	50	0.05 - 10 ppm	Biquinoline
9	pH*	486639	100	6.4 - 8.4 pH	Phenol Red
10	Total Iron, TPTZ (as Fe^{+2}/Fe^{+3})*	486650	50	0.02 - 4.5 ppm	TPTZ + PP

Included in the Standard Kit (486691-K)

**Combined Chlorine Test requires Free Chlorine (486637) to be run first.